


HV Board	Date	Guard voltage	Field voltage	Sense voltage	Current	Time	Super layer
HVB-1	3/20/2012	525	-700	1400	4uA	9:08	SL1
					1uA	11:30	
HVB-2	3/20/2012	525	-700	1400	8uA	9:30	SL1
					1uA	10:00	
HVB-3	3/21/2012	525	-700	1400	3uA	10:46	SL1
					1uA	11:00	
HVB-4	3/22/2012	525	-700	1100	20uA	9:30	SL1
				1100	.3uA	10:12	
				1200	9uA	10:13	
				1200	.6uA	10:14	
				1300	3uA	10:14	
				1400	16uA	10:21	
				1400	1.7uA	11:23	
HVB-5	3/23/2012	525	-700	1100	6uA	8:17	SL1
				1200	5uA	8:24	
				1300	13uA	8:29	
				1300	1.4uA	8:32	
				1400	5uA	8:32	
				1400	.9uA	8:36	
HVB-6	3/23/2012	525	-700	1100	15uA	9:08	SL1
				1100	1uA	9:09	
				1200	4uA	9:09	
				1300	16uA	9:11	
				1300	4uA	9:39	
				1400	25uA	9:45	
	3/26/2012			1400	.5uA	8:12	
HVB-7	3/26/2012	525	-700	1100	2uA	8:51	SL1
				1200	3uA	8:52	
				1300	5uA	8:53	
				1400	10uA	8:54	
	3/27/2012			1400	1.8uA	8:00	
HVB-1	4/6/2012	525	-700	1200	7uA	10:49	SL2
				1200	.2uA	11:03	
				1400	20uA	11:04	
				1400	.7uA	11:38	
HVB-2	4/9/2012	525	-700	1300	8uA	8:36	SL2
				1300	1.7uA	8:39	

				1400	11uA	8:39	
				1400	1.8uA	9:06	
HVB-3	4/9/2012	525	-700	1400	6uA	9:35	SL2
				1400	1.6uA	10:35	
HVB-4	4/9/2012	525	-700	1300	14uA	11:04	SL2
				1300	.9uA	11:08	
				1400	10uA	11:09	
				1400	1.9uA	11:41	
HVB-5	4/10/2012	525	-725	1000	16uA	9:15	SL2
				1000	.1uA	9:33	
				1200	14uA	9:35	
				1200	2uA	9:44	
				1300	20uA	9:44	
				1300	1uA	10:10	
				1400	22uA	10:10	
				1450	22uA	11:18	
	4/11/2012			1450	4uA	8:30	
HVB-6	4/11/2012	525	-725	825	3uA	9:48	SL2
				825	0uA	9:50	
				925	4uA	9:51	
				925	0uA	9:53	
				1125	5uA	9:54	
				1125	0uA	9:59	
				1325	6uA	10:02	
				1325	1.6uA	11:10	
				1400	25uA	11:19	
				1425	11uA	11:47	
				1450	25uA	12:00	
	4/12/2012			1450	2uA	8:00	
HVB-7	4/12/2012	525	-725	925	4uA	10:00	SL2
				925	0uA	10:03	
				1025		10:03	
		525	-700	1225	36uA	11:00	
				1225	2.6uA	11:34	
				1260	5uA	11:38	
				1260	2uA	11:40	
				1400	28uA	11:50	
	4/13/2012			1400	3uA	8:00	
				1450	3uA	8:01	



Increase voltage so we could see signal

Kept tripping so I turned field Voltage down